

CERTIFICATE OF CONFORMITY

Product: Laboratory Equipment

Name and address of the applicant: Break-A-Beam Incorporated
25257 Mound Road
Warren, MI 48091

Contact: Mr. James Nichols
Same as Applicant

Rating and principal characteristics: 120 V~, 60 Hz, 35mA / 12-30Vdc, 35mA

Trade Mark (if any):

Model: M-200/M-200-DC

Additional information (if necessary): Subject to the conditions of acceptability of the
Certification Report.

A sample of the product was tested and found
to be in conformity with: See Page 2

as shown in the Certification Report No: 9906171681
which forms part of this certificate

Products in Compliance with the Certification Report are
eligible to bear the Entela Certification Mark.

This Certificate is issued by:

Entela Inc.
Grand Rapids, MI, USA

Signature


Ned Devine
Program Manager III

Date 10 February 2000



Entela, Inc.
Engineering and Testing Laboratories
1-800-888-3787
www.entela.com

Page 2 of 2
Report No. 9906171681
10 February 2000
Break-A-Beam, Inc.

A Sample of the product was tested and found to be in conformity with:

UL 3111-1, Safety Electrical Measuring and Test Equipment, First Edition including revisions through 30 June 1994

CAN/CSA C22.2 No. 1010.1, Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use, First Edition including revisions through February 1997

UL 508, Safety Industrial Control Equipment, Seventeenth Edition including revisions through 16 July 1999

CSA C22.2 No. 205-M1983, Signal Equipment General Instruction, No. 1, R (1992)

FACILITIES:

- 3033 Madison Ave. SE • Grand Rapids, MI 49548 • Ph: (616) 247-0515 • Fax: (616) 247-7527
- 35550 Industrial Road • Livonia, MI 48150 • Ph: (734) 591-9161 • Fax: (734) 591-9349
- 81 Kelfield St., Unit 7 • Toronto, Ontario, Canada M9W 5A3 • Ph: (416) 241-8427 • Fax: (416) 241-0682